

Brian Aker

172

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-21-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2645	35
2 Bend Over	2271	41
3 Shake Head	5201	17
4 Redon 1	2645	35
5 Redon 2	4101	22
Test Result:	3063	30

PASS
Robert Frey
Brian Aker

Christoher Ameli

227

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-21-20**

Other: .

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	27382	3
2 Bend Over	10120	9
3 Shake Head	27382	3
4 Redon 1	9218	10
5 Redon 2	11354	8
Test Result:	13572	6

PASS
Robert Frey
Christoher Ameli

Stephen Ameli

214

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	1225	76
2 Bend Over	1446	64
3 Shake Head	3754	24
4 Redon 1	2369	39
5 Redon 2	3474	26
Test Result:	2013	46

PASS
Robert Frey
Stephen Ameli

Frank Balestriere Jr.

217

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **06-02-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	1263	73
2 Bend Over	3083	30
3 Shake Head	4330	21
4 Redon 1	1547	60
5 Redon 2	4138	22
Test Result:	2237	41

PASS
Robert Frey
Frank Balestriere Jr.

David Bassett

242

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Small Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-21-20**

Other:

Serial No: **1867**

Notes:

Step Number:	Fit Factor:	Leak Rate:
1 Face Forward	5201	17
2 Bend Over	2965	31
3 Shake Head	4476	20
4 Redon 1	7162	13
5 Redon 2	10120	9
Test Result:	5043	18

PASS
Robert Frey
David Bassett

Chris J Benderoth

188

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

Step Number:	Fit Factor:	Leak Rate:
1 Face Forward	640	145
2 Bend Over	2489	37
3 Shake Head	2206	42
4 Redon 1	2424	38
5 Redon 2	2155	43
Test Result:	1518	61

PASS
Robert Frey
Chris J Benderoth

James R Bette

1695

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

Step Number:	Fit Factor:	Leak Rate:
1 Face Forward	1730	53
2 Bend Over	1763	52
3 Shake Head	1981	47
4 Redon 1	503	185
5 Redon 2	2558	36
Test Result:	1241	75

PASS
Robert Frey
James R Bette

Ben Boursiquot

236

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2489	37
2 Bend Over	3895	23
3 Shake Head	1256	74
4 Redon 1	1099	84
5 Redon 2	3244	28
Test Result:	1871	49

PASS
Robert Frey
Ben Boursiquot

Gary Braus Jr.

248

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2699	34
2 Bend Over	5144	18
3 Shake Head	2551	36
4 Redon 1	3895	23
5 Redon 2	1699	54
Test Result:	2774	33

PASS
Robert Frey
Gary Braus Jr.

Philip Bussey

216

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-14-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2126	43
2 Bend Over	1708	54
3 Shake Head	1170	79
4 Redon 1	1651	56
5 Redon 2	2974	31
Test Result:	1753	53

PASS
Robert Frey
Philip Bussey

Christopher Carnovale

165

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-14-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	1952	47
2 Bend Over	3945	23
3 Shake Head	2126	43
4 Redon 1	4310	21
5 Redon 2	3073	30
Test Result:	2788	33

PASS
Robert Frey
Christopher Carnovale

Daniel Carrion

232

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-14-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	10231	9
2 Bend Over	2444	38
3 Shake Head	10231	9
4 Redon 1	3651	25
5 Redon 2	5855	15
Test Result:	4765	19

PASS
Robert Frey
Daniel Carrion

John Casiero

246

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-14-20**

Other:

Serial No: **1867**

Notes:

Step Number:	Fit Factor:	Leak Rate:
1 Face Forward	8464	11
2 Bend Over	15016	6
3 Shake Head	11494	8
4 Redon 1	2222	41
5 Redon 2	6698	13
Test Result:	5740	16

PASS
Robert Frey
John Casiero

Giuseppe Chirico

219

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **06-02-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	1322	70
2 Bend Over	2476	37
3 Shake Head	3023	30
4 Redon 1	3879	24
5 Redon 2	1485	62
Test Result:	2064	45

PASS
Robert Frey
Giuseppe Chirico

Ryan Ciccaglione

255

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-14-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2804	33
2 Bend Over	1994	46
3 Shake Head	2078	44
4 Redon 1	2327	40
5 Redon 2	4498	20
Test Result:	2511	37

PASS
Robert Frey
Ryan Ciccaglione

Michael J Connolly

185

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

Step Number:	Fit Factor:	Leak Rate:
1 Face Forward	4874	19
2 Bend Over	2064	45
3 Shake Head	5747	16
4 Redon 1	3474	26
5 Redon 2	3609	25
Test Result:	3500	26

PASS
Robert Frey
Michael J Connolly

Steven Coponi

137

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-21-20**

Other:

Serial No: **1867**

Notes:

Step Number:	Fit Factor:	Leak Rate:
1 Face Forward	2322	40
2 Bend Over	3062	30
3 Shake Head	8387	11
4 Redon 1	4702	19
5 Redon 2	1408	66
Test Result:	2779	33

PASS
Robert Frey
Steven Coponi

Glen Coppola

251

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-14-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	1297	71
2 Bend Over	3651	25
3 Shake Head	2327	40
4 Redon 1	1369	68
5 Redon 2	1597	58
Test Result:	1766	52

PASS
Robert Frey
Glen Coppola

Brendan Curry

262

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

Step Number:	Fit Factor:	Leak Rate:
1 Face Forward	2489	37
2 Bend Over	1363	68
3 Shake Head	2106	44
4 Redon 1	2064	45
5 Redon 2	1029	90
Test Result:	1630	57

PASS
Robert Frey
Brendan Curry

Steven Cusano

204

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-14-20**

Other:

Serial No: **1867**

Notes:

Step Number:	Fit Factor:	Leak Rate:
1 Face Forward	1952	47
2 Bend Over	1497	62
3 Shake Head	4119	22
4 Redon 1	3398	27
5 Redon 2	12931	7
Test Result:	2786	33

PASS
Robert Frey
Steven Cusano

Bruce D'Amato Jr.

241

Department:

Med Eval:

Job Code:

Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	1256	74
2 Bend Over	2699	34
3 Shake Head	1124	82
4 Redon 1	14323	6
5 Redon 2	3739	24
Test Result:	2089	44

PASS
Robert Frey
Bruce D'Amato Jr.

Kevin Dean

226

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **06-04-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	3278	28
2 Bend Over	668	139
3 Shake Head	2645	35
4 Redon 1	1740	53
5 Redon 2	2974	31
Test Result:	1618	57

PASS
Robert Frey
Kevin Dean

Joe R DeFinis

2057

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **04-26-19**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **01-07-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2754	33
2 Bend Over	13691	6
3 Shake Head	9598	9
4 Redon 1	5060	18
5 Redon 2	4586	20
Test Result:	5230	17

PASS
Robert Frey
Joe R DeFinis

Joe R DeFinis

2057

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	8167	11
2 Bend Over	2699	34
3 Shake Head	5413	17
4 Redon 1	3739	24
5 Redon 2	3739	24
Test Result:	4123	22

PASS
Robert Frey
Joe R DeFinis

Joseph DeLuise

229

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	5747	16
2 Bend Over	7569	12
3 Shake Head	2699	34
4 Redon 1	7569	12
5 Redon 2	14323	6
Test Result:	5691	16

PASS
Robert Frey
Joseph DeLuise

Joseph G DePasquale

256

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	1291	72
2 Bend Over	3135	29
3 Shake Head	7000	13
4 Redon 1	3042	30
5 Redon 2	2254	41
Test Result:	2489	37

PASS
Robert Frey
Joseph G DePasquale

Matthew J DeSacia

4567

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-14-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	4952	18
2 Bend Over	3800	24
3 Shake Head	4952	18
4 Redon 1	2645	35
5 Redon 2	3651	25
Test Result:	3791	24

PASS
Robert Frey
Matthew J DeSacia

John P Dillon

221

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2623	35
2 Bend Over	7569	12
3 Shake Head	3609	25
4 Redon 1	8239	11
5 Redon 2	4874	19
Test Result:	4476	20

PASS
Robert Frey
John P Dillon

Jason C Fregans

220

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2369	39
2 Bend Over	2623	35
3 Shake Head	5444	17
4 Redon 1	1536	60
5 Redon 2	2206	42
Test Result:	2391	38

PASS
Robert Frey
Jason C Fregans

Gabriel Gatto

239

Department:

Med Eval:

Job Code:

Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-14-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	4119	22
2 Bend Over	3278	28
3 Shake Head	9218	10
4 Redon 1	21651	4
5 Redon 2	2170	42
Test Result:	4298	21

PASS
Robert Frey
Gabriel Gatto

Gabriel Gatto

239

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Small Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **06-02-20**

Other:

Serial No: **1867**

Notes:

Step Number:	Fit Factor:	Leak Rate:
1 Face Forward	553	168
2 Bend Over	1371	67
3 Shake Head	678	137
4 Redon 1	629	148
5 Redon 2	633	147
Test Result:	696	133

PASS
Robert Frey
Gabriel Gatto

Sean Gaynor

195

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-14-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	1332	69
2 Bend Over	3651	25
3 Shake Head	3398	27
4 Redon 1	1804	51
5 Redon 2	4726	19
Test Result:	2398	38

PASS
Robert Frey
Sean Gaynor

John Gibson

210

Department:



Med Eval:

Job Code:

Mask: **Scott AV-3000HT Medium Full Face**

1.50	Protocol:	SCBA
03-02-20		93.10
500		06-02-20
	Serial No:	1867

Step Number:	Fit Factor:	Leak Rate:
1 Face Forward	3665	25
2 Bend Over	10344	9
3 Shake Head	2891	32
4 Redon 1	3410	27
5 Redon 2	2276	40
Test Result:	3453	26

PASS
Robert Frey
John Gibson

Nick C Gillo

224

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	1866	49
2 Bend Over	4655	20
3 Shake Head	14323	6
4 Redon 1	4433	21
5 Redon 2	1642	56
Test Result:	3021	30

PASS
Robert Frey
Nick C Gillo

Jonathan Gonzalez

231

Department:

Med Eval:

Job Code:

Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2310	40
2 Bend Over	8167	11
3 Shake Head	1511	61
4 Redon 1	1940	48
5 Redon 2	1326	70
Test Result:	2011	46

PASS
Robert Frey
Jonathan Gonzalez

Omar W Guzman

206

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	3244	28
2 Bend Over	2106	44
3 Shake Head	5747	16
4 Redon 1	3474	26
5 Redon 2	2424	38
Test Result:	3017	30

PASS
Robert Frey
Omar W Guzman

Sean Hector

223

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-14-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	4498	20
2 Bend Over	7162	13
3 Shake Head	3398	27
4 Redon 1	11494	8
5 Redon 2	15016	6
Test Result:	6174	15

PASS
Robert Frey
Sean Hector

George Hesse

202

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	3349	27
2 Bend Over	8952	10
3 Shake Head	4433	21
4 Redon 1	2363	39
5 Redon 2	5413	17
Test Result:	4020	23

PASS
Robert Frey
George Hesse

Thomas Heys

199

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Large Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-21-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	3637	25
2 Bend Over	5509	16
3 Shake Head	2645	35
4 Redon 1	2073	44
5 Redon 2	4952	18
Test Result:	3292	28

PASS
Robert Frey
Thomas Heys

Steven Hywel

108

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **06-02-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	1177	79
2 Bend Over	5712	16
3 Shake Head	12250	7
4 Redon 1	6465	14
5 Redon 2	3337	27
Test Result:	3204	29

PASS
Robert Frey
Steven Hywel

Jonathan M Jenkins**244**

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	3609	25
2 Bend Over	4433	21
3 Shake Head	11083	8
4 Redon 1	4433	21
5 Redon 2	8239	11
Test Result:	5320	17

PASS
Robert Frey
Jonathan M Jenkins

Douglas Johnson

215

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-21-20**

Other:

Serial No: **1867**

Notes:

Step Number:	Fit Factor:	Leak Rate:
1 Face Forward	1214	76
2 Bend Over	722	128
3 Shake Head	1277	72
4 Redon 1	734	126
5 Redon 2	5201	17
Test Result:	1100	84

PASS
Robert Frey
Douglas Johnson

Phillip A Johnson

237

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	3895	23
2 Bend Over	1900	49
3 Shake Head	2106	44
4 Redon 1	4874	19
5 Redon 2	1424	65
Test Result:	2309	40

PASS
Robert Frey
Phillip A Johnson

Daniel Keenan

263

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-21-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	3928	23
2 Bend Over	1408	66
3 Shake Head	2444	38
4 Redon 1	3513	26
5 Redon 2	2503	37
Test Result:	2429	38

PASS
Robert Frey
Daniel Keenan

Matthew Kehoe

249

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-21-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	5201	17
2 Bend Over	2965	31
3 Shake Head	2271	41
4 Redon 1	4101	22
5 Redon 2	3385	27
Test Result:	3313	28

PASS
Robert Frey
Matthew Kehoe

William Kelly

225

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2064	45
2 Bend Over	1760	52
3 Shake Head	4232	22
4 Redon 1	695	133
5 Redon 2	1614	57
Test Result:	1494	62

PASS
Robert Frey
William Kelly

Anthony Loreto

257

Department:

Med Eval:

Job Code:

Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	903	103
2 Bend Over	2623	35
3 Shake Head	2946	31
4 Redon 1	2856	32
5 Redon 2	14323	6
Test Result:	2224	41

PASS
Robert Frey
Anthony Loreto

Nicholas Manzella

258

Department:

Med Eval:

Job Code:

Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2201	42
2 Bend Over	2946	31
3 Shake Head	3135	29
4 Redon 1	3739	24
5 Redon 2	798	116
Test Result:	1899	49

PASS
Robert Frey
Nicholas Manzella

Joseph M Matarazzo

245

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	3135	29
2 Bend Over	1363	68
3 Shake Head	3609	25
4 Redon 1	1940	48
5 Redon 2	1642	56
Test Result:	2037	45

PASS
Robert Frey
Joseph M Matarazzo

Joseph Matullo

212

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-21-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	12931	7
2 Bend Over	5509	16
3 Shake Head	5509	16
4 Redon 1	5509	16
5 Redon 2	4702	19
Test Result:	5991	15

PASS
Robert Frey
Joseph Matullo

William McGee

155

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2064	45
2 Bend Over	5747	16
3 Shake Head	644	144
4 Redon 1	2489	37
5 Redon 2	9904	9
Test Result:	1842	50

PASS
Robert Frey
William McGee

Patrick McGovern

211

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	5747	16
2 Bend Over	5144	18
3 Shake Head	6510	14
4 Redon 1	9904	9
5 Redon 2	12413	7
Test Result:	7107	13

PASS
Robert Frey
Patrick McGovern

John McManus

197

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-21-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	1544	60
2 Bend Over	2796	33
3 Shake Head	2572	36
4 Redon 1	2503	37
5 Redon 2	2217	42
Test Result:	2227	41

PASS
Robert Frey
John McManus

Ryan McPherson

j

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	786	118
2 Bend Over	1794	51
3 Shake Head	1562	59
4 Redon 1	2424	38
5 Redon 2	3474	26
Test Result:	1577	59

PASS
Robert Frey
Ryan McPherson

Stephen L Milano

269

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	3349	27
2 Bend Over	1225	76
3 Shake Head	1699	54
4 Redon 1	986	94
5 Redon 2	3895	23
Test Result:	1681	55

PASS
Robert Frey
Stephen L Milano

Richard Mulligan

209

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-21-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2217	42
2 Bend Over	1708	54
3 Shake Head	6207	15
4 Redon 1	3385	27
5 Redon 2	1948	47
Test Result:	2492	37

PASS
Robert Frey
Richard Mulligan

Joseph Neto

243

Department:

Med Eval:

Job Code:

Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	3349	27
2 Bend Over	3135	29
3 Shake Head	8167	11
4 Redon 1	1699	54
5 Redon 2	4433	21
Test Result:	3217	28

PASS
Robert Frey
Joseph Neto

Fethnat C Ni Shuileabhain**4798**

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Small Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-21-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	1115	83
2 Bend Over	553	168
3 Shake Head	3167	29
4 Redon 1	4290	21
5 Redon 2	1948	47
Test Result:	1327	70

PASS
Robert Frey**Fethnat C Ni Shuileabhain**

Frank R Noborine

233

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	1326	70
2 Bend Over	3135	29
3 Shake Head	2201	42
4 Redon 1	1699	54
5 Redon 2	1402	66
Test Result:	1767	52

PASS
Robert Frey
Frank R Noborine

Donald G Peters

181

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	1699	54
2 Bend Over	2310	40
3 Shake Head	2856	32
4 Redon 1	1940	48
5 Redon 2	1940	48
Test Result:	2081	44

PASS
Robert Frey
Donald G Peters

Thomas Quigley

173

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	3135	29
2 Bend Over	2155	43
3 Shake Head	3349	27
4 Redon 1	3739	24
5 Redon 2	2771	33
Test Result:	2924	31

PASS
Robert Frey
Thomas Quigley

Scott Rechner

187

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **06-02-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2393	38
2 Bend Over	7217	12
3 Shake Head	3083	30
4 Redon 1	6746	13
5 Redon 2	3962	23
Test Result:	3902	23

PASS
Robert Frey
Scott Rechner

Michael P Reilly

222

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Small Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	8952	10
2 Bend Over	6556	14
3 Shake Head	1446	64
4 Redon 1	8239	11
5 Redon 2	4066	22
Test Result:	3779	24

PASS
Robert Frey
Michael P Reilly

Josh D Riley

250

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	1124	82
2 Bend Over	5144	18
3 Shake Head	648	143
4 Redon 1	2155	43
5 Redon 2	3474	26
Test Result:	1480	62

PASS
Robert Frey
Josh D Riley

Sergio Rivas

247

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-21-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	1090	85
2 Bend Over	751	124
3 Shake Head	2965	31
4 Redon 1	2073	44
5 Redon 2	6650	14
Test Result:	1553	59

PASS
Robert Frey
Sergio Rivas

Carmine Ruta

218

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-21-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2165	43
2 Bend Over	1836	50
3 Shake Head	12931	7
4 Redon 1	2271	41
5 Redon 2	3385	27
Test Result:	2748	33

PASS
Robert Frey
Carmine Ruta

Robert Ryan

196

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-21-20**


Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	5509	16
2 Bend Over	2271	41
3 Shake Head	17566	5
4 Redon 1	4101	22
5 Redon 2	3267	28
Test Result:	4069	22

PASS

Robert Frey

Robert Ryan

Leon Sederwall

254

Department:

Med Eval:

Job Code:

Mask: **Scott AV-3000HT Medium-Large Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-14-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	4726	19
2 Bend Over	5855	15
3 Shake Head	3945	23
4 Redon 1	1681	55
5 Redon 2	5230	17
Test Result:	3516	26

PASS
Robert Frey
Leon Sederwall

Daniel P Shelley

193

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	6125	15
2 Bend Over	4232	22
3 Shake Head	7569	12
4 Redon 1	2551	36
5 Redon 2	3474	26
Test Result:	4127	22

PASS
Robert Frey
Daniel P Shelley

Daniel R Shelley

260

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-21-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	1262	73
2 Bend Over	7162	13
3 Shake Head	2882	32
4 Redon 1	3385	27
5 Redon 2	14778	6
Test Result:	3045	30

PASS
Robert Frey
Daniel R Shelley

Kevin Skokandich

261

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2424	38
2 Bend Over	5444	17
3 Shake Head	12413	7
4 Redon 1	4655	20
5 Redon 2	2489	37
Test Result:	3866	24

PASS
Robert Frey
Kevin Skokandich

Rich C Smith

205

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	3754	24
2 Bend Over	4655	20
3 Shake Head	5144	18
4 Redon 1	1326	70
5 Redon 2	2558	36
Test Result:	2746	33

PASS
Robert Frey
Rich C Smith

Albert Souza

192

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **06-02-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	718	129
2 Bend Over	659	141
3 Shake Head	576	161
4 Redon 1	515	180
5 Redon 2	583	159
Test Result:	602	154

PASS
Robert Frey
Albert Souza

Albert Souza

192

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Small Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **06-02-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	526	176
2 Bend Over	587	158
3 Shake Head	576	161
4 Redon 1	629	148
5 Redon 2	580	160
Test Result:	578	161

PASS
Robert Frey
Albert Souza

Damon Stewart**182**

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-14-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	3167	29
2 Bend Over	3278	28
3 Shake Head	5230	17
4 Redon 1	4498	20
5 Redon 2	5542	16
Test Result:	4116	22

PASS
Robert Frey
Damon Stewart

Matthew Taylor

240

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-14-20**

Other:

Serial No: **1867**

Notes:

Step Number:	Fit Factor	Leak Rate:
1 Face Forward	1497	62
2 Bend Over	2722	34
3 Shake Head	3513	26
4 Redon 1	1708	54
5 Redon 2	3513	26
Test Result:	2283	40

PASS
Robert Frey
Matthew Taylor

Michael Taylor

191

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2106	44
2 Bend Over	1668	55
3 Shake Head	2771	33
4 Redon 1	2699	34
5 Redon 2	7000	13
Test Result:	2566	36

PASS
Robert Frey
Michael Taylor

Angelo Tedesco Jr.

176

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **06-04-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2509	37
2 Bend Over	2974	31
3 Shake Head	2387	39
4 Redon 1	7162	13
5 Redon 2	1279	72
Test Result:	2409	38

PASS
Robert Frey
Angelo Tedesco Jr.

Anthony O Torrissi

6280

Department:

Med Eval:

Job Code:

Mask: **Scott AV-3000HT Small Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

Step Number:	Fit Factor:	Leak Rate:
1 Face Forward	2201	42
2 Bend Over	3349	27
3 Shake Head	5747	16
4 Redon 1	4433	21
5 Redon 2	12413	7
Test Result:	4055	22

PASS
Robert Frey
Anthony O Torrissi

Mike P Triano

144

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	3474	26
2 Bend Over	921	101
3 Shake Head	11083	8
4 Redon 1	1829	50
5 Redon 2	3474	26
Test Result:	2175	42

PASS
Robert Frey
Mike P Triano

James Turco

228

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-14-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	10231	9
2 Bend Over	2387	39
3 Shake Head	4952	18
4 Redon 1	7162	13
5 Redon 2	7162	13
Test Result:	5011	18

PASS
Robert Frey
James Turco

James Turco

228

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Small Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **06-02-20**

Other:

Serial No: **1867**

Notes:

Step Number:	Fit Factor:	Leak Rate:
1 Face Forward	746	124
2 Bend Over	629	148
3 Shake Head	751	123
4 Redon 1	734	126
5 Redon 2	576	161
Test Result:	679	137

PASS
Robert Frey
James Turco

Joseph Urso

235

Department:

Med Eval:

Job Code:

Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **06-02-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2249	41
2 Bend Over	3337	27
3 Shake Head	2015	46
4 Redon 1	4213	22
5 Redon 2	2476	37
Test Result:	2657	35

PASS
Robert Frey
Joseph Urso

Craig Vanderhoof

238

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-14-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2387	39
2 Bend Over	2444	38
3 Shake Head	3398	27
4 Redon 1	1262	73
5 Redon 2	5855	15
Test Result:	2397	38

PASS
Robert Frey
Craig Vanderhoof

Heidi Vanderhoof

201

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-14-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	2722	34
2 Bend Over	924	100
3 Shake Head	1651	56
4 Redon 1	1994	46
5 Redon 2	2804	33
Test Result:	1716	54

PASS
Robert Frey
Heidi Vanderhoof

Anthony Vecchio

200

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-15-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	856	108
2 Bend Over	2771	33
3 Shake Head	1940	48
4 Redon 1	1124	82
5 Redon 2	887	105
Test Result:	1231	75

PASS
Robert Frey
Anthony Vecchio

William R Wagner

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	3042	30
2 Bend Over	2369	39
3 Shake Head	5144	18
4 Redon 1	7569	12
5 Redon 2	11083	8
Test Result:	4282	21

PASS
Robert Frey
William R Wagner

Paul H Wannemacher

167

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-20-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	1829	50
2 Bend Over	5144	18
3 Shake Head	3895	23
4 Redon 1	7000	13
5 Redon 2	2155	43
Test Result:	3116	29

PASS
Robert Frey
Paul H Wannemacher

Anthony V Williams

8467

Department:

Med Eval:

Job Code:


Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-21-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	687	135
2 Bend Over	1912	48
3 Shake Head	2714	34
4 Redon 1	5855	15
5 Redon 2	9218	10
Test Result:	1904	48

PASS
Robert Frey
Anthony V Williams

Abdur Yasin

208

Department:

Med Eval:

Job Code:



Mask: **Scott AV-3000HT Medium Full Face**Challenge Pressure: **1.50**Protocol: **SCBA**Last NIST Calibration: **03-02-20**Respiratory Rate: **93.10**Minimum Passing Fit Factor: **500**Last Daily Calibration: **05-14-20**

Other:

Serial No: **1867**

Notes:

<u>Step Number:</u>	<u>Fit Factor:</u>	<u>Leak Rate:</u>
1 Face Forward	707	131
2 Bend Over	609	152
3 Shake Head	1079	86
4 Redon 1	2579	36
5 Redon 2	2271	41
Test Result:	1039	89

PASS
Robert Frey
Abdur Yasin